

# WP 410.03

Set of 10 ISO-V specimens 10x10, CuZn



#### Description

### ISO-V specimens as accessories for the WP 410 experimental unit

This set of ISO-V specimens is available as an accessory for the WP 410 experimental unit. In a Charpy test, a notched specimen whose two ends lie on a support is broken or bent by hammer impact and drawn through the supports. The resistance of the specimen against this load provides an insight into the notched bar impact strength of the material. The results of the Charpy test offer a reference for the properties of the material. However, they are only valid for comparison when they are attained with geometrically equivalent specimens under the same test conditions. For this reason, only ISO profile specimens are used as well as notched specimens with a rectangular profile (GUNT- profile).

The set contains ten ISO-V specimens made of brass.

## Learning objectives/experiments

- together with WP 410
  - determine the notched-bar impact work
  - determine the notched-bar impact strength
  - analyse the fracture surface characteristics
  - plot a notched-bar impact work-temperature diagram

### Specification

- [1] specimens for Charpy notched-bar impact test
- [2] 10 specimens made of brass, form ISO-V
- [3] accessory for WP 410 Impact test, 300Nm

# Technical data

10 notched-bar impact specimens

- form: ISO-V
- LxWxH: 55x10x10mm
- notch radius: 0,25mm
- material: brass

Weight: approx. 500g

### Scope of delivery

1 set of specimens (10 pieces)



# WP 410.03 Set of 10 ISO-V specimens 10x10, CuZn

Required accessories

WP 410 Impact test, 300Nm

G.U.N.T. Gerätebau GmbH, Hanskampring 15-17, D-22885 Barsbüttel, Telefon (040) 67 08 54-0, Fax (040) 67 08 54-42, Email sales@gunt.de, Web www.gunt.de We reserve the right to modify our products without any notifications. Page 2/2 - 11.2023